



CODA-CERVA

VETERINARY AND AGROCHEMICAL RESEARCH CENTRE

GROESELBERG 99 – B 1180 BRUSSELS (UKKEL)

TEL: +32 (0)2 379 04 11

FAX : + 32 (0)2 379 06 70

HTTP: // WWW.CODA-CERVA.BE



172-PT

PROFICIENCY TESTING 2012

Enzootic Bovine Leukosis (EBL)

***Detection of EBL-specific antibodies in bovine serum by
Enzyme Linked Immunosorbent Assay (ELISA)***

OPERATIONAL UNIT

COORDINATION OF VETERINARY DIAGNOSIS

EPIDEMIOLOGY AND RISK ASSESSMENT

(CVD-ERA)

DATE BEGIN PT: 29 OCTOBER 2012

DATE REPORT: 27 NOVEMBER 2012

I. Introduction

Details relevant to the proficiency test (PT) are available in the Procedure PRO/2.5/01 'Beheer van de proficiency testen op het CODA-CERVA-Ukkel/Gestion des essais d'aptitude au CODA-CERVA-Uccle', which is summarized in the 'Manual for the participant'.

II. Aim

The aim of this PT was to evaluate the ability of the participating laboratories to identify the absence or presence of EBL-specific antibodies in individual bovine serum by ELISA.

III. Materials and methods

III.1. Conduct of diagnostic tests

In the framework of this PT, predefined reference serum samples must be tested by means of an EBL antibody ELISA test. The procedures for the ELISA tests must be fully described in the SOPs of the participating laboratories.

III.2. Reference samples

Replicates of 5 reference serum samples of bovine origin, either free from detectable EBL-specific antibodies (n=2; coded 'PT2012EBLSERNS1' and 'PT2012EBLSERNS2') or containing detectable EBL-specific antibodies (n=3; coded 'PT2012EBLSERPS1', 'PT2012EBLSERPS2' and 'PT2012EBLSERPS3'), were used. In total, 120 aliquots were distributed to 6 participating laboratories. All participants were given 4 aliquots of the 5 reference serum samples, i.e. 20 aliquots. The positions of the reference serum samples in the sent blocks were randomized for each participant (Table 3).

For each reference serum sample, a certificate containing the status of the sample (= 'golden standard') was made. The status of the reference serum samples was based on (i) the historical background of the animals and (ii) the results obtained by an immunodiffusion assay and the SERELISA BLV Ab Mono Blocking ELISA kit from Synbiotics (batch 11 SBLV1 99) (pre-verification). The reference serum samples PT2012EBLSERNS1 and PT2012EBLSERNS2 were obtained from EBL-free animals. In contrast, the reference serum samples PT2012EBLSERPS1 and PT2012EBLSERPS3 were derived from animals that were experimentally infected with EBL, whereas the reference serum sample PT2012EBLSERPS2 was a 1/2 dilution of a serum obtained from an animal that was EBL antibody positive upon birth and stayed EBL antibody positive during its entire life. For each reference serum sample, the same qualitative result was obtained with both test methods used. Taken together, the reference serum samples PT2012EBLSERNS1 and PT2012EBLSERNS2 were considered as negative sera, and the reference serum samples PT2012EBLSERPS1, PT2012EBLSERPS2 and PT2012EBLSERPS3 as variably positive sera in EBL antibody ELISA.

After aliquoting the different reference serum samples, a homogeneity check was performed on 10 aliquots of each reference serum sample using the SERELISA BLV Ab Mono Blocking ELISA kit from Synbiotics (batch 11 SBLV1 100), hereby obtaining the same qualitative result for all 10 aliquots of the same reference serum sample. Consequently, all reference serum samples were considered as reliable samples in order to evaluate the ability of laboratories to correctly identify the absence or presence of EBL-specific antibodies in bovine serum. In addition, all reference serum samples were tested once after the PT in order to confirm their stability and status (post-verification) using the SERELISA BLV Ab Mono Blocking ELISA kit from Synbiotics (batch 11 SBLV1 102).

III.3. Classification of results, level of agreement and threshold for qualification

III.3.1. Classification of results

Results provided by the participating laboratories are categorized as *success* (positive result when the reference sample is truly positive, negative result when the reference sample is truly negative) or *failure* (positive result when the reference sample is truly negative, negative result when the reference sample is truly positive, non-interpretable result when the reference sample is truly negative or positive).

III.3.2. Level of agreement

The level of agreement achieved by the participating laboratories is expressed as the percentage of *success* (i.e., the reported result matches with the assigned status) for the 20 aliquots of reference serum samples used for this PT.

III.3.3. Threshold for qualification

Following the procedure, a participating laboratory is only qualified if the level of agreement for the 20 aliquots of reference serum samples is at least 90%.

IV. Results

For confidentiality reasons, the participating laboratories are quoted anonymously and the concordance table is safely kept at the operational unit CVD-ERA of CODA-CERVA.

IV.1. Transfer and start of the analyses of the reference samples

The 20 aliquots of reference serum samples were sent frozen (dry ice) to each of the 6 participating laboratories by national or international courier on 29th of October 2012 (120 aliquots in total). LAB3, LAB4, LAB5 and LAB6 acknowledged receipt of the samples on the same day, whereas LAB2 and LAB1 acknowledged receipt of the samples on 30th and 31st of October 2012, respectively. Analyses were performed between 30th of October and 8th of November 2012 (Table 1).

IV.2. Dates at which results were returned to the operational unit CVD-ERA

Results were submitted to the operational unit CVD-ERA between 7th and 12th of November 2012. LAB4 hereby exceeded the deadline of 9th of November 2012 for submission of the results (Table 1).

Table 1. Overview of the dates on which (i) the reference serum samples were received and analyzed by the participating laboratories, and (ii) the obtained results were submitted to the operational unit CVD-ERA of CODA-CERVA.

Laboratory	Reference samples received	Start of analysis	Submission of the results (Excel file)
LAB1	31/10/2012	08/11/2012	09/11/2012
LAB2	30/10/2012	06/11/2012	09/11/2012
LAB3	29/10/2012	30/10/2012	07/11/2012
LAB4	29/10/2012	05/11/2012	12/11/2012
LAB5	29/10/2012	07/11/2012	09/11/2012
LAB6	29/10/2012	07/11/2012	09/11/2012

IV.3. Compliance with the procedure

Except LAB4, all participating laboratories have provided a duly dated and signed copy of the results.

IV.4. Qualitative data analysis

IV.4.1. Level of agreement

Qualitative data analysis showed that 5 out of 6 participating laboratories (LAB1 until LAB5) provided qualitative results that were in full agreement with the true status of the reference serum samples (100% of agreement), whereas LAB6 misclassified 1 aliquot and hence obtained 95% of agreement (Table 2).

A quantitative data analysis (including box plots) is shown for educational purposes in Annex 1 and Annex 2.

Table 2. Agreement between results generated by the participating laboratories (LABNR) and the status of the reference serum samples assigned by the EBL reference laboratory of CODA-CERVA. All participating laboratories received 20 aliquots of reference serum samples. Results are presented as absolute values and percentages (in parentheses).

	LABNR					
	1	2	3	4	5	6
failure	0 (0.0)	0 (0.0)	0 (0.0)	0 (0.0)	0 (0.0)	1 (5.0)
success	20 (100.0)	20 (100.0)	20 (100.0)	20 (100.0)	20 (100.0)	19 (95.0)

IV.4.2. Variability among participating laboratories

No variability in qualitative laboratory results could be observed between LAB1 until LAB5 since these participants correctly identified all reference serum samples. In contrast, LAB6 misclassified 1 aliquot of the negative reference serum sample PT2012EBLSERNS2 (POS instead of NEG).

For each participating laboratory, the obtained results and the assigned statuses for the reference serum samples are shown in Table 3.

Table 3. The responses (RESULT) of the participating laboratories (LABNR) with the identification of the reference serum samples (SAMPLE), the positions of the reference serum samples as placed in the block (LABPOSIT), and the status assigned by the EBL reference laboratory of CODA-CERVA (STATUS). NEG: negative; POS: positive.

	LABNR	LABPOSIT	SAMPLE	STATUS	RESULT	SUCCESS
1	1	1	PT2012EBLSERNS1	NEG	NEG	1
2	1	2	PT2012EBLSERPS1	POS	POS	1
3	1	3	PT2012EBLSERNS1	NEG	NEG	1
4	1	4	PT2012EBLSERPS1	POS	POS	1
5	1	5	PT2012EBLSERNS1	NEG	NEG	1
6	1	6	PT2012EBLSERPS2	POS	POS	1
7	1	7	PT2012EBLSERNS1	NEG	NEG	1
8	1	8	PT2012EBLSERPS2	POS	POS	1
9	1	9	PT2012EBLSERNS2	NEG	NEG	1
10	1	10	PT2012EBLSERPS2	POS	POS	1
11	1	11	PT2012EBLSERNS2	NEG	NEG	1
12	1	12	PT2012EBLSERPS2	POS	POS	1
13	1	13	PT2012EBLSERNS2	NEG	NEG	1
14	1	14	PT2012EBLSERPS3	POS	POS	1
15	1	15	PT2012EBLSERNS2	NEG	NEG	1
16	1	16	PT2012EBLSERPS3	POS	POS	1
17	1	17	PT2012EBLSERPS1	POS	POS	1
18	1	18	PT2012EBLSERPS3	POS	POS	1
19	1	19	PT2012EBLSERPS1	POS	POS	1
20	1	20	PT2012EBLSERPS3	POS	POS	1
21	2	1	PT2012EBLSERNS1	NEG	NEG	1
22	2	2	PT2012EBLSERPS2	POS	POS	1
23	2	3	PT2012EBLSERNS1	NEG	NEG	1
24	2	4	PT2012EBLSERPS2	POS	POS	1
25	2	5	PT2012EBLSERNS2	NEG	NEG	1
26	2	6	PT2012EBLSERPS2	POS	POS	1
27	2	7	PT2012EBLSERNS2	NEG	NEG	1
28	2	8	PT2012EBLSERPS2	POS	POS	1
29	2	9	PT2012EBLSERNS2	NEG	NEG	1
30	2	10	PT2012EBLSERPS3	POS	POS	1
31	2	11	PT2012EBLSERNS2	NEG	NEG	1
32	2	12	PT2012EBLSERPS3	POS	POS	1
33	2	13	PT2012EBLSERPS1	POS	POS	1
34	2	14	PT2012EBLSERPS3	POS	POS	1
35	2	15	PT2012EBLSERPS1	POS	POS	1
36	2	16	PT2012EBLSERPS3	POS	POS	1
37	2	17	PT2012EBLSERNS1	NEG	NEG	1
38	2	18	PT2012EBLSERPS1	POS	POS	1
39	2	19	PT2012EBLSERNS1	NEG	NEG	1
40	2	20	PT2012EBLSERPS1	POS	POS	1



(Table 3 - CONTINUED)

	LABNR	LABPOSIT	SAMPLE	STATUS	RESULT	SUCCESS
41	3	1	PT2012EBLSERNS2	NEG	NEG	1
42	3	2	PT2012EBLSERPS2	POS	POS	1
43	3	3	PT2012EBLSERNS2	NEG	NEG	1
44	3	4	PT2012EBLSERPS2	POS	POS	1
45	3	5	PT2012EBLSERNS2	NEG	NEG	1
46	3	6	PT2012EBLSERPS3	POS	POS	1
47	3	7	PT2012EBLSERNS2	NEG	NEG	1
48	3	8	PT2012EBLSERPS3	POS	POS	1
49	3	9	PT2012EBLSERPS1	POS	POS	1
50	3	10	PT2012EBLSERPS3	POS	POS	1
51	3	11	PT2012EBLSERPS1	POS	POS	1
52	3	12	PT2012EBLSERPS3	POS	POS	1
53	3	13	PT2012EBLSERNS1	NEG	NEG	1
54	3	14	PT2012EBLSERPS1	POS	POS	1
55	3	15	PT2012EBLSERNS1	NEG	NEG	1
56	3	16	PT2012EBLSERPS1	POS	POS	1
57	3	17	PT2012EBLSERNS1	NEG	NEG	1
58	3	18	PT2012EBLSERPS2	POS	POS	1
59	3	19	PT2012EBLSERNS1	NEG	NEG	1
60	3	20	PT2012EBLSERPS2	POS	POS	1
61	4	1	PT2012EBLSERNS2	NEG	NEG	1
62	4	2	PT2012EBLSERPS3	POS	POS	1
63	4	3	PT2012EBLSERNS2	NEG	NEG	1
64	4	4	PT2012EBLSERPS3	POS	POS	1
65	4	5	PT2012EBLSERPS1	POS	POS	1
66	4	6	PT2012EBLSERPS3	POS	POS	1
67	4	7	PT2012EBLSERPS1	POS	POS	1
68	4	8	PT2012EBLSERPS3	POS	POS	1
69	4	9	PT2012EBLSERNS1	NEG	NEG	1
70	4	10	PT2012EBLSERPS1	POS	POS	1
71	4	11	PT2012EBLSERNS1	NEG	NEG	1
72	4	12	PT2012EBLSERPS1	POS	POS	1
73	4	13	PT2012EBLSERNS1	NEG	NEG	1
74	4	14	PT2012EBLSERPS2	POS	POS	1
75	4	15	PT2012EBLSERNS1	NEG	NEG	1
76	4	16	PT2012EBLSERPS2	POS	POS	1
77	4	17	PT2012EBLSERNS2	NEG	NEG	1
78	4	18	PT2012EBLSERPS2	POS	POS	1
79	4	19	PT2012EBLSERNS2	NEG	NEG	1
80	4	20	PT2012EBLSERPS2	POS	POS	1

(Table 3 - CONTINUED)

	LABNR	LABPOSIT	SAMPLE	STATUS	RESULT	SUCCESS
81	5	1	PT2012EBLSERPS1	POS	POS	1
82	5	2	PT2012EBLSERPS3	POS	POS	1
83	5	3	PT2012EBLSERPS1	POS	POS	1
84	5	4	PT2012EBLSERPS3	POS	POS	1
85	5	5	PT2012EBLSERNS1	NEG	NEG	1
86	5	6	PT2012EBLSERPS1	POS	POS	1
87	5	7	PT2012EBLSERNS1	NEG	NEG	1
88	5	8	PT2012EBLSERPS1	POS	POS	1
89	5	9	PT2012EBLSERNS1	NEG	NEG	1
90	5	10	PT2012EBLSERPS2	POS	POS	1
91	5	11	PT2012EBLSERNS1	NEG	NEG	1
92	5	12	PT2012EBLSERPS2	POS	POS	1
93	5	13	PT2012EBLSERNS2	NEG	NEG	1
94	5	14	PT2012EBLSERPS2	POS	POS	1
95	5	15	PT2012EBLSERNS2	NEG	NEG	1
96	5	16	PT2012EBLSERPS2	POS	POS	1
97	5	17	PT2012EBLSERNS2	NEG	NEG	1
98	5	18	PT2012EBLSERPS3	POS	POS	1
99	5	19	PT2012EBLSERNS2	NEG	NEG	1
100	5	20	PT2012EBLSERPS3	POS	POS	1
101	6	1	PT2012EBLSERNS1	NEG	NEG	1
102	6	2	PT2012EBLSERPS1	POS	POS	1
103	6	3	PT2012EBLSERNS1	NEG	NEG	1
104	6	4	PT2012EBLSERPS1	POS	POS	1
105	6	5	PT2012EBLSERNS1	NEG	NEG	1
106	6	6	PT2012EBLSERPS2	POS	POS	1
107	6	7	PT2012EBLSERNS1	NEG	NEG	1
108	6	8	PT2012EBLSERPS2	POS	POS	1
109	6	9	PT2012EBLSERNS2	NEG	POS	0
110	6	10	PT2012EBLSERPS2	POS	POS	1
111	6	11	PT2012EBLSERNS2	NEG	NEG	1
112	6	12	PT2012EBLSERPS2	POS	POS	1
113	6	13	PT2012EBLSERNS2	NEG	NEG	1
114	6	14	PT2012EBLSERPS3	POS	POS	1
115	6	15	PT2012EBLSERNS2	NEG	NEG	1
116	6	16	PT2012EBLSERPS3	POS	POS	1
117	6	17	PT2012EBLSERPS1	POS	POS	1
118	6	18	PT2012EBLSERPS3	POS	POS	1
119	6	19	PT2012EBLSERPS1	POS	POS	1
120	6	20	PT2012EBLSERPS3	POS	POS	1

V. Discussion

The purpose of this PT was to assess the performances of the participating laboratories when analyzing individual reference serum samples of bovine origin for the detection of EBL-specific antibodies by ELISA.

Based on qualitative data analysis, 5 out of 6 participating laboratories (LAB1 until LAB5) provided results that were in full agreement with the true status of the reference serum samples (100% of agreement), whereas LAB6 misclassified 1 aliquot of the negative reference serum sample PT2012EBLSERNS2 (95% of agreement) (Table 2).

EBL antibody ELISA kits from 2 different producers as well as different batches from the same producer were used: Synbiotics (2 batches: 11 SBLV1 100 and 11 SBLV1 102) and IDEXX Montpellier SAS (1 batch: 1211). LAB1 used the EBL antibody ELISA kit from IDEXX Montpellier SAS, whereas the other participants used the EBL antibody ELISA kit from Synbiotics. Hereby, LAB3, LAB4, LAB5 and LAB6 used the same batch (11 SBLV1 100). In addition, at least 3 of these laboratories performed the long incubation protocol for the conjugate (LAB3, LAB4 and LAB6; LAB5 did not provide information about the used incubation protocol).

VI. Conclusions

According to the procedure currently in force, the performance of a participating laboratory is satisfactory if at least 90% of the results provided by this laboratory is in agreement with the status of the reference serum samples assigned by the EBL reference laboratory of CODA-CERVA (see III.3.3.). Consequently, all participants achieved a satisfactory performance for the detection of EBL-specific antibodies in reference serum samples by ELISA.

Head CVD-ERA
Yves Van der Stede

Appendix

Name of the participating laboratories

Agence nationale de sécurité sanitaire de l'alimentation, de l'environnement et du travail (ANSES) (Niort, France)

Association Régionale de Santé et d'Identification Animales (ARSIA) (Ciney, Belgium)

Dierengezondheidszorg Vlaanderen (DGZ) (Torhout, Belgium)

Laboratoire de Médecine Vétérinaire de l'Etat (LMVE) (Grand Duchy of Luxemburg)

Synbiotics Europe (Lyon, France)

Veterinary and Agrochemical Research Center (CODA-CERVA) (Ukkel, Belgium)



Annex 1: Quantitative data analysis

Besides qualitative data analysis (positive, negative or non-interpretable result), also quantitative data analysis was performed using the statistical software programs SAS 9.2 (summary statistics) and R (box plots). All quantitative data analyses were performed on normalized data, namely the percentages blocking calculated according to the instructions of the PT provider: $[(\text{mean OD}_{\text{Negative Kit Controls}} - \text{OD}_{\text{Sample}}) / (\text{mean OD}_{\text{Negative Kit Controls}} - \text{mean OD}_{\text{Positive Kit Controls}})] * 100$.

The quantitative data analysis in this report was not used to evaluate the participants in this PT, but should only be considered as educational information for the participants in order to evaluate their performance and/or to standardize their different diagnostic tests.

I. Box plots

Box plots of the percentages blocking per reference serum sample and per participating laboratory were made using the statistical software R and are shown in Figure 1.

Remark: To calculate the percentages blocking, the PT provider used the formula corresponding with the manual of the EBL antibody ELISA kit from Synbiotics. Because the percentages blocking according to the manual of the EBL antibody ELISA kit from IDEXX Montpellier SAS are calculated using a slightly different formula, namely $[(\text{mean OD}_{\text{Negative Kit Controls}} - \text{OD}_{\text{Sample}}) / (\text{mean OD}_{\text{Negative Kit Controls}})] * 100$, the cut-off for the EBL antibody ELISA kit from IDEXX Montpellier SAS was adapted accordingly (64% instead of 60%).




Figure 1. Box plots showing the percentage blocking per reference sample serum sample and per participating laboratory. Box plots represent the minimum value, the maximum value, the median, the lower (25%) and upper (75%) quartile, and possible outliers per

sample and per laboratory. (Adapted) cut-off values for the EBL antibody ELISA kit from IDEXX Montpellier SAS (64%) and Synbiotics (30-50%) are shown in green and red, respectively. LAB1 used the EBL antibody ELISA kit from IDEXX Montpellier SAS, whereas the other participants used the EBL antibody ELISA kit from Synbiotics. Hereby, LAB3, LAB4, LAB5 and LAB6 used the same batch and at least LAB3, LAB4 and LAB6 performed the same incubation protocol for the conjugate.

II. Mandel's h- and k-statistics (z-scores)

Based on ISO 5725-2 and ISO 13528, between-lab variability (reproducibility) and within-lab variability (repeatability) were estimated through Mandel's h- and k-statistics, respectively, using the statistical software SAS 9.2. Mandel's h- and k-statistics were calculated based on percentages blocking per reference serum sample and per participating laboratory.

The h-statistic depends on the number of participants, whereas the k-statistic depends on both the number of participants and the number of repeats per sample. When 30 participants or more are involved in a PT, a satisfactory between-lab and within-lab consistency is obtained when the (absolute) value for the h- and k-statistic is smaller than 2. An unsatisfactory result (a corrective action is required) is reached when the (absolute) value is larger than 3. (Absolute) values between 2 and 3 indicate a questionable consistency. Importantly, in case of a smaller number of participants (which is the case in this PT), other indicator values apply for Mandel's h- and k-statistics (Table 1).

Table 1. Indicators for Mandel's h- and k-statistics at the 5% significance level in function of the number of participating laboratories (p) and the number of repeats per sample (n) as described in ISO 5725-2.

p (# labs)	h	k								
		n (# repeats)								
		2	3	4	5	6	7	8	9	10
3	1,15	1,65	1,53	1,45	1,40	1,37	1,34	1,32	1,30	1,29
4	1,42	1,76	1,59	1,50	1,44	1,40	1,37	1,35	1,33	1,31
5	1,57	1,81	1,62	1,53	1,46	1,42	1,39	1,36	1,34	1,32
6	1,66	1,85	1,64	1,54	1,48	1,43	1,40	1,37	1,35	1,33
7	1,71	1,87	1,66	1,55	1,49	1,44	1,41	1,38	1,36	1,34
8	1,75	1,88	1,67	1,56	1,50	1,45	1,41	1,38	1,36	1,34
9	1,78	1,90	1,68	1,57	1,50	1,45	1,42	1,39	1,36	1,35
10	1,80	1,90	1,68	1,57	1,50	1,46	1,42	1,39	1,37	1,35

Based on Table 1, the maximum absolute value for Mandel's h-statistic for this PT is 1,66 (p=6), whereas the maximum value for Mandel's k-statistic is 1,54 for all reference serum samples (p=6 and n=4).

LAB2, LAB3, LAB4 and LAB5 obtained a satisfactory between-laboratory consistency for all reference serum samples, whereas the other participants showed increased values for Mandel's h-statistic for 1 reference serum sample: LAB1 for PT2012EBLSERPS1 (k=1,83) and LAB6 for PT2012EBLSERNS2 (k=1,91).

LAB1, LAB2 and LAB4 obtained a satisfactory within-laboratory consistency for all reference serum samples, whereas the other participants showed increased values for Mandel's k-statistic for at least 1 reference serum sample: LAB3 for PT2012EBLSERNS1 (k=1,69) and PT2012EBLSERPS1 (k=2,15), LAB5 for PT2012EBLSERPS2 (k=1,56) and PT2012EBLSERPS3 (k=1,78), and LAB6 for PT2012EBLSERNS2 (k=1,96).

All data used for the calculations of Mandel's h- and k-statistics can be found in Annex 2.

III. ANOVA

Using a SAS macro encoding a general linear model (GLM) with laboratories as fixed effect and the normalized OD values (in this case the percentages blocking) as a dependent variable, it was investigated whether statistically significant differences exist ($\alpha=0,05$) between participating laboratories. Comparisons were made at the global level (all reference serum samples were analysed together), status level (all reference serum samples with the same status were analysed together) and sample level (all reference serum samples were analysed individually). Since comparing quantitative results



between participants or methods (e.g. different kits, batches or incubation protocols) is most relevant at the status level (less variation than at a global level), we focused on the latter.

No statistically significant differences were observed between laboratories at a global level. However, statistically significant differences existed at both sample and status level. At the status level, significant differences were observed for both the negative and positive reference serum samples. LAB6 reported percentages blocking that were significantly higher than those reported by LAB2 for the negative reference serum samples, whereas LAB1 reported percentages blocking that were significantly higher than those reported by LAB2 and LAB4 for the positive reference serum samples.



Annex 2: Calculations of Mandel's h- and k-statistics (based on % blocking)

Sample	Labnr	n_i	v_i	x_i_m	x_g_m	between_lab_coeff	STDEV_repeat	STDEV_repro	STDEV_betweenlab	h	k	cv
PT2012EBLSERNS1	1	4	38,93	3,46	-7,36	0,14	8,98	9,68	3,60	1,17	0,69	180,08
PT2012EBLSERNS1	2	4	54,66	-21,03	-7,36	0,14	8,98	9,68	3,60	-1,48	0,82	-35,15
PT2012EBLSERNS1	3	4	230,04	-14,50	-7,36	0,14	8,98	9,68	3,60	-0,77	1,69	-104,60
PT2012EBLSERNS1	4	4	5,22	-8,95	-7,36	0,14	8,98	9,68	3,60	-0,17	0,25	-25,52
PT2012EBLSERNS1	5	4	102,28	-1,43	-7,36	0,14	8,98	9,68	3,60	0,64	1,13	-708,22
PT2012EBLSERNS1	6	4	52,93	-1,70	-7,36	0,14	8,98	9,68	3,60	0,61	0,81	-427,02
PT2012EBLSERNS2	1	4	9,17	-0,04	5,36	0,09	10,12	10,62	3,21	-0,61	0,30	-6729,36
PT2012EBLSERNS2	2	4	65,36	-2,34	5,36	0,09	10,12	10,62	3,21	-0,88	0,80	-345,36
PT2012EBLSERNS2	3	4	45,74	1,67	5,36	0,09	10,12	10,62	3,21	-0,42	0,67	405,78
PT2012EBLSERNS2	4	4	15,62	6,51	5,36	0,09	10,12	10,62	3,21	0,13	0,39	60,75
PT2012EBLSERNS2	5	4	86,75	4,22	5,36	0,09	10,12	10,62	3,21	-0,13	0,92	220,87
PT2012EBLSERNS2	6	4	391,68	22,13	5,36	0,09	10,12	10,62	3,21	1,91	1,96	89,44
PT2012EBLSERPS1	1	4	0,02	101,02	99,88	0,04	0,94	0,96	0,18	1,83	0,14	0,13
PT2012EBLSERPS1	2	4	0,30	99,74	99,88	0,04	0,94	0,96	0,18	-0,23	0,58	0,55
PT2012EBLSERPS1	3	4	4,13	99,17	99,88	0,04	0,94	0,96	0,18	-1,15	2,15	2,05
PT2012EBLSERPS1	4	4	0,14	99,90	99,88	0,04	0,94	0,96	0,18	0,02	0,40	0,38
PT2012EBLSERPS1	5	4	0,63	99,91	99,88	0,04	0,94	0,96	0,18	0,05	0,84	0,80
PT2012EBLSERPS1	6	4	0,13	99,56	99,88	0,04	0,94	0,96	0,18	-0,52	0,38	0,36
PT2012EBLSERPS2	1	4	0,10	95,13	86,09	0,68	2,68	4,73	3,89	1,03	0,12	0,33
PT2012EBLSERPS2	2	4	15,23	71,46	86,09	0,68	2,68	4,73	3,89	-1,66	1,46	5,46
PT2012EBLSERPS2	3	4	2,54	92,00	86,09	0,68	2,68	4,73	3,89	0,67	0,59	1,73
PT2012EBLSERPS2	4	4	3,65	80,28	86,09	0,68	2,68	4,73	3,89	-0,66	0,71	2,38
PT2012EBLSERPS2	5	4	17,46	86,81	86,09	0,68	2,68	4,73	3,89	0,08	1,56	4,81
PT2012EBLSERPS2	6	4	4,10	90,85	86,09	0,68	2,68	4,73	3,89	0,54	0,76	2,23
PT2012EBLSERPS3	1	4	0,02	100,61	69,93	0,93	2,55	9,77	9,43	1,45	0,05	0,14
PT2012EBLSERPS3	2	4	1,88	64,37	69,93	0,93	2,55	9,77	9,43	-0,26	0,54	2,13



Sample	Labnr	n_i	v_i	x_i_m	x_g_m	between_lab_coeff	STDEV_repeat	STDEV_repro	STDEV_betweenlab	h	k	cv
PT2012EBLSERPS3	3	4	1,33	92,33	69,93	0,93	2,55	9,77	9,43	1,06	0,45	1,25
PT2012EBLSERPS3	4	4	12,24	54,64	69,93	0,93	2,55	9,77	9,43	-0,72	1,37	6,40
PT2012EBLSERPS3	<u>5</u>	4	20,65	54,84	69,93	0,93	2,55	9,77	9,43	-0,71	1,78	8,29
PT2012EBLSERPS3	6	4	3,04	52,80	69,93	0,93	2,55	9,77	9,43	-0,81	0,68	3,30

Legend: Labnr = number attributed to a laboratory during the PT; n_i = number of replicates; v_i = total variability (variance) in the normalised data (% blocking); x_i_m = mean of normalized data (% blocking); x_g_m = mean of normalized data (% blocking) obtained by all laboratories; between_lab_coeff = fraction of total variability due to differences between labs for each sample; STDEV_repeat = repeatability standard deviation over all laboratories; STDEV_repro = reproducibility standard deviation over all laboratories; STDEV_betweenlab = between-lab standard deviation over all laboratories; h-statistic = between-laboratory consistency; k-statistic = within-laboratory consistency; CV = variation coefficient in %. Values for Mandel's h- and k-statistics shown in red/underlined/bold exceed the corresponding limit value as determined in Annex 1 (Table 1).